

NPI 442

OXIDATION STABILITY OF GASOLINE

(INDUCTION PERIOD METHOD - DRY BATH)



STANDARDS

ASTM D 525, D 873, IP 40, IP 138, ISO 7536 and related methods.

SCOPE

This test method covers the determination of the stability of gasoline in finished form only, under accelerated oxidation conditions. The induction period may be used as an indication of the tendency of motor gasoline to form gum in storage. It should be recognized, however, that its correlation with the formation of gum in storage may vary markedly under different storage conditions and with different gasolines.



NORMALAB
www.normalab.com

SPECIFICATIONS

The NPI 442, dry bath version, offers graphic screen with friendly software. It can record the last 200 results. The temperature range goes from ambient to 120°C. The device covers the ASTM D525 and D 873 standards.

Key features

- 4 positions
- **Test Temperature Range: Ambient to 120°C**
- Computerized and fully automated test
- **Direct pressure measurement in the vessel**
- Quick coupling

Software

- Possibility to monitor the test up to 96 hours
- Data Storage: 200 results
- Print test result and graph
- Residue method

Accessories

- REF 40930 **NPC 210** Safety oxygen filling and decompression unit with protective screen



Ready to be connected to the ground
Weight : 30 kg
Dimension : 500 x 365 x 800mm

- REF 513513 Glass test container with cover
- REF 40903/1 Stainless steel test vessel with safety burst disc



ORDERING INFORMATION

40925

Scope of delivery:

NPI 442 is delivered with:

- 4 Pressure sensors (P/N 40914)

According to your application, accessories and options must be ordered separately.

Site requirements:

- Power supply: AC 230 V, 50 Hz - 10 A
- Dimension: 570x 630x 640 mm
- Weight: 40 kg

Oxygen extra-dry 99,6% purity

Reagents: gums solvent (mixture of toluene and acetone 99%)

 **NORMALAB**
www.normalab.com

CONTACT : sales@normalab.com

Normalab FRANCE SAS
ZA Caux Multipôles 1 - 76190 Valliquerville
Tel. : +33 232.700.100
Fax : +33 232.704.732

DISTRIBUTED BY